

# ***VacStation***

Extraorális elszívó rendszer



KIFINOMULT MÉRNÖKI MUNKA

ÁTLÁTSZÓ ELSZÍVÓTÁNYÉR

KÜLÖNLEGES SZÉNKEFEMENTES MOTOR

HAJLÍTHATÓ FLEXIBILIS CSŐ



## A VACSTATION NÉGY SZINTŰ VÉDELMI RENDSZERE:

HIGH FIBER

GLASS FIBER  
HEPA 13

HEPA 13  
CERAMSIT + AKTÍV SZÉN  
+ KMNO4

SZÉNKEFEMENTES MOTOR

UV FÉNY

1

HighFiber légtechnikai vatta szűrő, amely kiszűri az aerosolt és a nedvességet (1 hónap élettartam, a készülékhez 6 db jár)

2

GlassFiber HEPA 13 szűrő, amely a port és egyéb lebegő részeket nyeli el 99%-os hatásfokkal (3-6 hónap élettartam)

3

Ceramsit + Aktív Szén + KMNO4 + HEPA H13 szűrő, ami megköt, szűr és sterilizál (6-12 hónap élettartam)

4

1210 uw/cm2 hatóerejű 280 nm hullámhosszon dolgozó UV sterilizáló

# ÁRLISTA



VacStation - 297.509 HUF/ db

HighFiber vatta szűrő - 1.665 HUF/db

Glass Fiber HEPA 13 szűrő - 15.938 HUF/db


Ceramsit HEPA H13 szűrő - 60.564 HUF/db

*MINDEN VÁSÁRLÁS  
MELLÉ AJÁNDÉK  
2 DB HIGHFIBER  
SZÜRŐ!*

## SZÁLLÍTÁSI IDŐ:

hajón szállítva (kb. 2 hónap),  
DHL-el repülően (kb. 1 hónap)  
(+ 113.792 HUF)

A gyártó csak a rendelkezésre álló készlettelől függően tudja kielégíteni a rendeléseket.



A KÉSZÜLÉKET SZŰRŐKKEL FELSZERELVE SZÁLLÍTJUK!

## EC Declaration of Conformity

Manufacturer: **Changzhou Sifary Medical Technology Co., Ltd.**  
**No. 99, Qingyang Road, Xuejia County, Xinbei District, Changzhou**  
**City, Jiangsu Province, China 213000**  
**Tel/Fax: 0519-85962691**

Authorized European Representative: **Llins Service & Consulting GmbH**  
**Obere Seegasse 34/2, 69124, Heidelberg, Germany**  
**Tel: +49 175 4870819**

Product name: **Extraoral Dental Vacuum System**  
Model Number: **VacStation**

Classification (MDD, Annex IX): **Class I, rule 12**

Conformity Assessment Route: **Annex VII**

We herewith declare that the above mentioned products meet the transposition into national law, the provisions of the following EC Council Directives and Standards. All supporting documentations are retained under the premises of the manufacturer.

### DIRECTIVES

#### General applicable directives:

Medical Device Directive: COUNCIL DIRECTIVE 93/42/EEC of 14 June 1993 concerning medical devices (MDD 93/42/EEC), as amended by Directive 2007/47/EC.

Standard Applied: All applied harmonized standards.

Notified Body: N/A

NB Identification number: N/A

(EC) Certificate(s): N/A

Expire date of the Certificate: N/A

Place, Date of Issue: **Changzhou, 2020-04-10**

Signature: 

Name: Tao Huan

Position: General Manager